

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 22D2126206	(X3) Date Survey Completed 12/04/2018
Name of Provider or Supplier Haven Medical Group	Street Address, City, State 86 Baker Ave Extension, Concord, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A CLIA recertification survey was conducted for the Haven Medical Group laboratory pursuant to the Clinical Laboratory Improvement Amendments (CLIA) of 1988 and CLIA regulations at 42 CFR 493.
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to perform calibration verifications as appropriate as evidenced by the following: Quality control (QC)</p>

record review for calendar year 2018 revealed that calibration verifications of at least 3 points were not performed every six months for three (3) out of three (3) routine chemistry analytes on the Thermo Fisher Indiko. Calibration verifications were last performed on 8/28/17. The laboratory director interviewed on 12/4/18 at 3:30 PM confirmed that calibration verifications of at least 3 points had not been performed at least once every six months. The laboratory performs a total of 6,240 routine chemistry tests annually.