

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 22D2133197	(X3) Date Survey Completed 03/08/2024
Name of Provider or Supplier Precision Toxicology Llc DbA Precision Diagnostics	Street Address, City, State 49 State Road, Nauset Bldg, Suite 103, Dartmouth, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of the laboratory work area and interview with the General Supervisor (GS), the laboratory failed to establish a safety procedure to protect employees from hazardous waste in the laboratory work area. Findings include: 1. On 2/26/2024 at 2:00 PM a 55 gallon drum containing liquid chemical waste from high performance liquid chromatography (HPLC) composed of water, ethanol, methanol, ammonium acetate and formic acid was observed in the laboratory accessioning and initial drug screening work area. 2. Record review on 2/29/2024 of an email received on 2/28/2024 from the GS in reference to the contents of the 55 gallon drum noted in #1, the GS stated, "It's mostly HPLC liquid waste which consists of water, some methanol and very small amounts of formic acid." 3. Record review on 3/8/2024 of an email received on 3/7/2024 from the GS in reference to the contents of the 55 gallon drum noted in #1, the GS stated, "There could be sometimes very small amounts of ethanol or ammonium acetate in addition to the ones described before in the hazardous waste barrel. The hazardous waste sticker on the side of the barrel has the waste classified as "Flammable Liquid", EPA Waste Code D001." 4. Surveyor review on 3/8/2024 of the Environmental Protection Agency website's definition of EPA Waste Code D001 revealed, "EPA assigned D001 as the waste code for ignitable hazardous wastes." 5. Record review on 3/8/2024 of the laboratory's 'EHS 102 Hazard Communication Written Program,' procedure attached to an email received on 3/9/2024 from the GS revealed, the procedure did not address the safe storage of 55 gallons of chemical waste in laboratory work areas. 6. Record review on 3/8/2024 of</p>

the laboratory's 'EHS 105 Safety Policy and Training Program,' procedure attached to an email received on 3/9/2024 from the GS revealed, the procedure did not address the safe storage of 55 gallons of chemical waste in laboratory work areas. 7. Staff interview on 2/26/2024 with the GS confirmed the findings in #1 above. The GS stated, "We do not have anywhere else to store this 55 gallon drum of chemical waste."