

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 22D2159614	<b>(X3) Date Survey Completed</b> 10/16/2025
<b>Name of Provider or Supplier</b> Cgt Global	<b>Street Address, City, State</b> 1230 Soldiers Field Rd, Suite 2, Boston, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A CLIA recertification survey was conducted for the CGT Global laboratory on 10/16 /2025 pursuant to the Clinical Laboratory Improvement Amendments (CLIA) of 1988 and CLIA regulations at 42 CFR 493.
<b>D2128</b>	<p>HEMATOLOGY CFR(s): 493.851(e)</p> <p>(e)(1) For any unsatisfactory analyte or test performance or testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure. (2) For any unacceptable analyte or testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event.</p> <p>This STANDARD is not met as evidenced by:  . Based on record review of the laboratory's College of American Pathologists (CAP) proficiency testing (PT) scores and interview with the Technical Consultant (TC), the laboratory failed to investigate and perform corrective action when a PT score of less than 100% was received. The surveyor reviewed the CAP PT results for calendar years 2024 events A, B, and C and 2025 events A and B on 10/16/2025. The review revealed the laboratory received a score of 60% for the analyte Hematocrit and a score of 0% for the WBC differential for the CAP Event FH1-B-2025 Hematology Automated Differential Survey. The PT Performance Evaluation report was signed and reviewed by the Laboratory Director (LD), but the laboratory did not investigate or document corrective action for the 60% score for the Hematocrit and the 0% score for the WBC differential. The TC confirmed in an interview on 10/16/2025 at 11:38 AM that the laboratory did not investigate or document corrective action for the 60%</p>

score for the Hematocrit and the 0% score for the WBC differential. The TC stated, "The laboratory performed corrective action but did not document it." The laboratory performs 2,280 Hematology tests annually.