

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 22D2244033	(X3) Date Survey Completed 05/28/2026
Name of Provider or Supplier American Health Ne, Llc	Street Address, City, State 171 Forbes Blvd, Suite 2000, Mansfield, MA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A CLIA recertification survey was conducted for the American Health NE, LLC laboratory on 05/28/2026 pursuant to the Clinical Laboratory Improvement Amendments (CLIA) of 1988 and CLIA regulations at 42 CFR 493.
D5805	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>(c) The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the Laboratory Director (LD) and Technical Supervisor (TS) on 05/28/2026, the laboratory failed to indicate on the patient final test report the correct name of the laboratory where the test was performed as evidenced by the following: The surveyor reviewed ten (10) patient final test reports between September 2024 and April 2026. The review revealed that the laboratory failed to indicate the correct name of the laboratory where the test was performed for ten (10) out of ten (10) patient final test reports reviewed. The name of the laboratory on the patient's final test reports was American Health Associates. The laboratory's name on the CLIA certificate is American Health NE, LLC. The LD and TS confirmed in an interview on 05/28/2026 at 2:49 P.M. that the patient final test reports did not indicate the correct name of the laboratory where the test was performed. The laboratory performs 370 Bacteriology, 42 Immunology, 908,780 Chemistry, and 315,048 Hematology tests annually.</p>