

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  23D0362833	<b>(X3) Date Survey Completed</b>  04/21/2026
<b>Name of Provider or Supplier</b>  Bloom Pediatrics	<b>Street Address, City, State</b>  2055 E 14 Mile Road, Birmingham, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	. A proficiency testing desk review survey was performed on April 21, 2026, by the State of Michigan Department of Licensing and Regulatory Affairs. The laboratory was found out of compliance with CLIA regulations (42 CFR Part 493, Laboratory Requirements) for the following Condition(s): 493.803 Condition: Successful participation
<b>D2016</b>	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: . Based on record review of the CMS database, the American Association of Bioanalysts - Medical Laboratory Evaluation (AAB - MLE) proficiency testing scorecards and College of American Pathologist (CAP) proficiency testing scorecards,</p>

the laboratory failed to achieve satisfactory performance for 2 (2nd Event 2025 and 1st Event 2026) of 3 testing events for the specialty Hematology for the analyte of Hematocrit (HCT). Refer to D2130.

**D2130**

**HEMATOLOGY**

CFR(s): 493.851(f)

(f) Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

. Based on record review of the CMS database, the American Association of Bioanalysts - Medical Laboratory Evaluation (AAB - MLE) proficiency testing scorecards and College of American Pathologist (CAP) proficiency testing scorecards, the laboratory failed to achieve satisfactory performance for 2 (2nd Event 2025 and 1st Event 2026) of 3 testing events for the specialty Hematology for the analyte of Hematocrit (HCT). Findings include: 1. A record review of the CMS database, American Association of AAB - MLE proficiency testing scorecards and College of CAP proficiency testing scorecards revealed the laboratory failed achieve satisfactory performance for 2 out of 3 consecutive proficiency testing (PT) events: Hematology Testing PT Event Score 2nd event 2025 60% (AAB-MLE) 1st event 2026 60% (CAP)