

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  23D0363649	<b>(X3) Date Survey Completed</b>  03/10/2022
<b>Name of Provider or Supplier</b>  Gladstone A Payton Do	<b>Street Address, City, State</b>  247 N Gratiot Ave, Mount Clemens, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5807</b>	<p>TEST REPORT CFR(s): 493.1291(d)</p> <p>Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.</p> <p>This STANDARD is not met as evidenced by:                      . Based on record review and interview with Technical Consultant #1 (TC1), the laboratory failed to include the reference intervals for testosterone testing for its female patients for 2 (March 2020 to March 2022) of 2 years reviewed. Findings include: 1. A review of patient test records revealed the reference intervals on test reports for testosterone were 190-890 pg/mL for both male and female patients. 2. An interview on 3/10/22 at 10:45 am with TC1 revealed the reference interval of 190-890 pg/mL is for male patients and that same reference interval is used on all patient test reports regardless of sex.</p>