

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0365804	(X3) Date Survey Completed 08/11/2020
Name of Provider or Supplier Forum Medical Clinics, Pc	Street Address, City, State 25625 Schoenherr Road, Warren, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: . Based on record review of the CMS database and the American Proficiency Institute (API) proficiency testing reports, it was determined the laboratory failed to successfully participate in a CMS-approved proficiency testing program for the specialty of chemistry analyte sodium. Findings include: Review of the CMS database and the API proficiency testing reports revealed the laboratory failed to achieve satisfactory performance for the specialty of chemistry analyte sodium. Refer to D2096. This is a repeat deficiency from 4/2/2019.</p>
D2096	ROUTINE CHEMISTRY

CFR(s): 493.841(f)

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

. Based on review of the CMS database and the American Proficiency Institute (API) proficiency testing reports, the laboratory failed to achieve satisfactory performance for the chemistry analyte sodium in 2 (1st event of 2020 and 2nd event of 2020) of 2 consecutive testing events. Findings include: 1. A review of the CMS database and the API proficiency testing reports revealed the laboratory had unsatisfactory performance for 2 of 2 consecutive proficiency testing events constituting unsuccessful performance for the chemistry analyte sodium: Sodium PT Event Score 1st event 2020 60% 2nd event 2020 60% This is a repeat deficiency from 4/2/2019.

D6000

MODERATE COMPLEXITY LABORATORY DIRECTOR

CFR(s): 493.1403

The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.

This CONDITION is not met as evidenced by:

. Based on review of the CMS database and the American Proficiency Institute (API) proficiency testing reports, the Laboratory Director failed to provide overall management and direction in accordance with 493.1407 of this subpart. Findings include: 1. The Laboratory Director failed to ensure the laboratory successfully participated in a proficiency testing program as required under Subpart H. Refer to D6016.

D6016

LABORATORY DIRECTOR RESPONSIBILITIES

CFR(s): 493.1407(e)(4)(i)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;

This STANDARD is not met as evidenced by:

. Based on record review of the CMS database and the American Proficiency Institute (API) proficiency testing reports, the Laboratory Director failed to ensure the laboratory successfully participated in a proficiency testing program as required under Subpart H, resulting in subsequent unsuccessful performance for 2 (1st event of 2020 and 2nd event of 2020) out of 2 consecutive testing events. Findings include: 1. A record review of the CMS database and API proficiency testing reports revealed subsequent unsatisfactory performance for 2 out of 2 consecutive testing events for the chemistry analyte sodium: Sodium PT Event Score 1st event 2020 60% 2nd event

2020 60% 2. A record review of the CMS database revealed initial unsuccessful proficiency testing performance for the chemistry analyte sodium: Sodium PT event Score 3rd event 2018 60% 1st event 2019 60%