

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0365804	(X3) Date Survey Completed 09/06/2022
Name of Provider or Supplier Forum Medical Clinics, Pc	Street Address, City, State 25625 Schoenherr Road, Warren, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: . Based on record review of the CMS database and the American Proficiency Institute (API) proficiency testing reports, it was determined the laboratory failed to successfully participate in a CMS-approved proficiency testing program for the chemistry analyte iron. Findings include: The laboratory failed to attain a score of at least 80% of acceptable responses for the chemistry analyte iron, which is unsatisfactory performance. Refer to D2087.</p>
D2087	<p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p>

Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.

This STANDARD is not met as evidenced by:

. Based on review of the CMS database and the American Proficiency Institute (API) proficiency testing reports, the laboratory failed to attain a score of at least 80% of acceptable responses for the chemistry analyte iron, which is unsatisfactory performance for 2 (2022 1st and 2nd events) of 2 testing events in 2022. Findings include: PT Event Iron Score 1st event 2022 60% 2nd event 2022 60%