

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0367552	(X3) Date Survey Completed 08/19/2019
Name of Provider or Supplier Michigan Healthcare Professionals P C	Street Address, City, State 2011 Middlebelt Road, Garden City, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5777	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(b)(c)</p> <p>(b) The laboratory must have a system to identify and assess patient test results that appear inconsistent with the following relevant criteria, when available: (b)(1) Patient age. (b)(2) Sex. (b)(3) Diagnosis or pertinent clinical data. (b)(4) Distribution of patient test results. (b)(5) Relationship with other test parameters. (c) The laboratory must document all test result comparison activities.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with testing personnel #1 (TP1), the laboratory failed to document the comparison studies for the sampling of the patient specimen in the open mode versus the closed mode methods on the Beckman Coulter Act diff 2 hematology analyzer for three (2nd six months in 2017 and 2018) of four twice a year comparisons. Findings included: 1. Record review revealed the laboratory lacked the documentation to show the comparison studies were performed and documented twice a year for the following: a. 2nd comparison in 2017 b. 1st and 2nd comparisons in 2018 2. During the interview on August 19, 2019 at 10:15 a.m., TP1 confirmed the laboratory failed to perform and document the comparison studies twice a year.</p>