

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0369899	(X3) Date Survey Completed 05/02/2019
Name of Provider or Supplier Louis C Chiara Md	Street Address, City, State 120 S Main, Suite D, Milford, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: . Based on observation, patient log review, and interview with the office manager, the laboratory was using mycobiotic media that exceeded the manufacturer's expiration date for 1 of 3 lots used from 2017 to 2019. Findings include: 1. An observation on 5/2 /19 at 9:18 am by the surveyor during the laboratory tour revealed a lot of Troy Biological Mycobiotic Agar (1805210) had expired on 2/21/29 and was in use. 2. A review of the patient testing log showed 11 cases had been tested since the current lot of mycobiotic media exceeded the 2/21/19 expiration date. 3. In an interview on 5/2 /19 at 9:18 am, the office manager acknowledged the mycobiotic media was expired.</p>