

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  23D0369930	<b>(X3) Date Survey Completed</b>  11/07/2022
<b>Name of Provider or Supplier</b>  Junction Clinic Pc	<b>Street Address, City, State</b>  4771 Michigan Avenue, Detroit, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D3011</b>	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by:                      . Based on observation, record review, and interview with the Medical Assistant, the laboratory failed to ensure safety procedures were established and observed to ensure safety from biohazardous materials by storing food in the laboratory refrigerator. Findings include: 1. The surveyor observed cheese wrapped in paper resting on hematology controls containing blood, jars of food, and a molding vegetable in the refrigerator on 11/7/22 at 10:19 am. 2. A review of the laboratory's procedures revealed a lack of safety procedures established. 3. An interview on 11/7/22 at 10:22 am with the Medical Assistant confirmed food to be present in the laboratory refrigerator.</p>