

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0370869	(X3) Date Survey Completed 02/01/2021
Name of Provider or Supplier Hamzavi Dermatology Grosse Pointe	Street Address, City, State 16815 E Jefferson Avenue Suite 140, Grosse Pointe, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5785	<p>CORRECTIVE ACTIONS CFR(s): 493.1282(b)(3)</p> <p>(b) The laboratory must document all corrective actions taken, including actions taken when any of the following occur: (b)(3) The criteria for proper storage of reagents and specimens, as specified under 493.1252(b), are not met.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the laboratory liaison, the laboratory failed to document corrective action for improper storage of the Michelles Medium, Viral Culture, and Fungal Culture media for 2 (February 2019 to February 2021) of two years reviewed. Findings include: 1. A record review of the "Room and Equipment Quality Control - Temperature Log Sheet" revealed for 2 of 2 years the refrigerator that stored the Michelles Medium, Viral Culture, and Fungal Culture media the temperature was outside of the stated range of 35.6 - 46.4 on specific days with no corrective action taken as follows: a. 7/15/19 - 46.5 b. 1/8/20 - 35.4 c. 6/8/20 - 34.5 d. 6/11/20 - 33.9 e. 6/22/20 - 34.8 f. 6/24/20 - 35.0 g. 6/25/20 - 34.6 h. 6/29/20 - 35.0 i. 7/2/20 - 35.0 j. 7/13/20 - 35.0 k. 7/20/20 - 35.4 l. 8/20/20 - 31.8 m. 8/24/20 - 35.2 n. 8/27/20 - 35.0 o. 11/3/20 - 35.0 p. 11/16/20 - 35.4 2. A interview on 2/01/21 at 10:30 am, the laboratory liaison confirmed that no corrective action was taken for the temperatures outside the stated range.</p>