

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0372910	(X3) Date Survey Completed 11/29/2022
Name of Provider or Supplier Skin & Vein Center	Street Address, City, State 305 N Leroy Street, Fenton, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the laboratory liaison (LL), the laboratory failed to verify the accuracy of its Potassium Hydroxide (KOH) and Scabies preparations at least twice annually for 1 (2022) of 2 years. Findings include: 1. A review of the laboratory's "KOH comparison" chart review revealed a lack of documentation for 1 of 2 years of testing. 2. When queried on 11/29/2022 at 12:27 pm, the LL stated that no specimens had been tested thus far in 2022 so there was nothing to compare. 3. An interview on 11/29/2022 at 12:27 pm, the LL confirmed the KOH and scabies twice annual verification of accuracy was not completed in 2022.</p>