Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0669482	(X3) Date Survey Completed 05/03/2022
Name of Provider or Supplier Together Women's Health	**Street Address, City, State** 39200 Garfield - B, Clinton Township, MI	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: . Based on review of the CMS database and the American Association of Bioanalysts (AAB) proficiency testing reports, it was determined that the laboratory failed to successfully participate in a CMS approved proficiency testing program for the bacteriology analyte: throat/urine culture identification. Findings include: Review of the CMS database and the AAB proficiency testing reports showed unsatisfactory performance for 2 of 2 proficiency testing events for the bacteriology analyte: throat /urine culture identification. Refer to D2028.
D2028	BACTERIOLOGY

	CFR(s): 493.823(e) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: . Based on review of the CMS database and review of the American Association of Bioanalysts (AAB) final proficiency testing program reports, the laboratory failed to achieve satisfactory performance for microbiology bacteriology throat/urine culture identification in 2 out of 2 consecutive testing events. Findings include: Throat/urine culture identification PT Event Score 1st event 2022 68% 3rd event 2021 71%