

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 23D0685754 | (X3) Date Survey Completed 02/14/2022 |
| Name of Provider or Supplier Pontiac General Hospital | Street Address, City, State 461 W Huron, Pontiac, MI | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|---|
| D2016 | <p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: . Based on record review of the CMS database and the American Proficiency Institute (API) proficiency testing reports, it was determined the laboratory failed to successfully participate in a CMS-approved proficiency testing program for immunohematology compatibility testing. Findings include: Review of the CMS database and the API proficiency testing reports revealed the laboratory failed to achieve satisfactory performance for compatibility testing. Refer to D2173 and D2181.</p> |
| D2181 | COMPATIBILITY TESTING |

CFR(s): 493.863(e)

Failure to achieve an overall testing event score of satisfactory for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

. Based on review of the CMS database and the American Proficiency Institute (API) proficiency testing reports, the laboratory failed to achieve satisfactory performance for 2 (3rd Event 2021 and 1st Event 2021) of 3 testing events for immunohematology compatibility testing. Findings include: Compatibility Testing PT Event Score 3rd event 2021 80% 1st event 2021 80%