Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0694147	(X3) Date Survey Completed 12/19/2024
Name of Provider or Supplier Psoriasis And Eczema Treatment Center	**Street Address, City, State** 597 Baldwin St, Jenison, MI	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5415	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c) Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (1) Identity and when significant, titer, strength or concentration. (2) Storage requirements. (3) Preparation and expiration dates. (4) Other pertinent information required for proper use. This STANDARD is not met as evidenced by: Based on observation and interview with the Medical Assistant (MA), the laboratory failed to label one of one aliquot of mineral oil used for scabies testing with preparation and expiration dates. Findings include: 1. The surveyor observed one aliquot of mineral oil in the laboratory on 12/19/2024 at 9:15 am without preparation or expiration date indicated on the bottle. 2. A laboratory policy titled "Quality Control", Section II, Page 18, "Reagents, Materials and Supplies," states, " Reagents.... must be labeled and must not be used when they have exceeded their expiration dates...". 3. An interview conducted on 12/19/2024 at 9:26 am with the MA confirmed that the aliquot was not labeled with a preparation or expiration date.
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: . Based on observation and interview with the Medical Assistant (MA), the laboratory

	failed to ensure that mineral oil used in scabies testing was not expired. Findings include: 1. The surveyor observed one of one bottle of mineral oil with an expiration date of 1/2018 in the laboratory on 12/19/2024 at 9:15 am. 2. An interview conducted on 12/19/2024 at 9:26 am with the MA confirmed the mineral oil was expired. 3. A laboratory policy titled "Quality Control", Section II, Page 18, "Reagents, Materials and Supplies," states, " Reagents....must be labeled and must not be used when they have exceeded their expiration dates...".