

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0701648	(X3) Date Survey Completed 06/10/2026
Name of Provider or Supplier Wsu Infectious Disease Laboratory	Street Address, City, State 275 E Hancock Street Room 039, Detroit, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	. A proficiency testing desk review survey was performed on June 10, 2026 by the Michigan Department of Licensing and Regulatory Affairs. The laboratory was found to be out of compliance with CLIA regulations (42 CFR Part 493, Laboratory Requirements) for the following Condition(s): 493.803 Condition: Successful participation
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of the CMS database, College of American Pathologists (CAP)

	proficiency testing reports, and interview with the laboratory manager, the laboratory failed to achieve satisfactory performance in the specialty of Microbiology for the test of Mycology. Refer to D2046.
D2046	MYCOLOGY CFR(s): 493.827(e) (e) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: . Based on review of the CMS database and College of American Pathologists proficiency testing reports, the laboratory failed to achieve satisfactory performance for 2 (2nd Event 2025 and 1st Event 2026) of 3 testing events in the specialty of Microbiology for the test of Mycology. Findings include: 1. A review of the CMS database and College of American Pathologists proficiency testing reports revealed the laboratory obtained the following proficiency testing scores for the test Mycology: PT Event Mycology Score 2nd Event 2025 Unsatisfactory 0 1st Event 2026 Unsatisfactory 0