Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0719175	(X3) Date Survey Completed 08/09/2018
Name of Provider or Supplier Farmington Dermatologists	**Street Address, City, State** 23133 Orchard Lake Road Suite 201, Farmington, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3043	RETENTION REQUIREMENTS CFR(s): 493.1105(a)(7) The laboratory must retain cytology slide preparations for at least 5 years from the date of examination (see 493.1274(f) for proficiency testing exception). The laboratory must retain histopathology slides for at least 10 years from the date of examination. The laboratory must retain pathology specimen blocks for at least 2 years from the date of examination. The laboratory must preserve remnants of tissue for pathology examination until a diagnosis is made on the specimen. This STANDARD is not met as evidenced by: . Based on record review and interview, the laboratory failed retain the histopathology slides for three (#3, #4, and #5) of nine patient charts audited for ten years. Findings include: 1. On August 9, 2018 at 10:26 AM, record review revealed for three (#3, #4, and #5) of nine patient charts audited the histopathology slide was not available to the surveyor. 2. During the interview on August 9, 2018 at 10:26 AM, the office staff confirmed the laboratory did not retain the histopathology slides for ten years.