

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0723829	(X3) Date Survey Completed 10/27/2021
Name of Provider or Supplier Michigan Institute Of Medicine, P C	Street Address, City, State 38525 8 Mile Road, Livonia, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on lack of documentation and interview with Testing Personnel #1 (TP1), the laboratory failed to review and evaluate the endocrinology results on the American Association of Bioanalysts (AAB) proficiency testing final graded report for one (event #3 of 2020) of seven events reviewed. Findings include: 1. A document review for the AAB proficiency testing reports revealed a lack of documentation of the final AAB proficiency testing reports for one (event #3 of 2020) of seven events performed. 2. When queried on 10/27/2021 at 11:55 am, TP1 was unable to provide the surveyor the final event report to show the results had been reviewed and evaluated. 3. An interview on 10/27/2021 at 11:55 am, TP1 confirmed the laboratory did not have documents of the final proficiency testing reports for event #3 of 2020, so no review was conducted.</p>