

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0879485	(X3) Date Survey Completed 12/17/2020
Name of Provider or Supplier Hamzavi Dermatology	Street Address, City, State 2950 Keewahdin Road, Fort Gratiot, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the Office Manager, the laboratory failed to verify the accuracy of its Potassium Hydroxide (KOH) and Scabies preparations for 2 (December 2018 to December 2020) of 2 years. Findings include: 1. A review of the laboratory's records revealed a lack of verification of accuracy testing for KOH and Scabies preparations between December 2018 and December 2020. 2. An interview on 12/17/20 at 1:31 pm with the Office Manager confirmed the laboratory did not perform verification of accuracy testing for KOH and Scabies preparations.</p>