

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D0945208	(X3) Date Survey Completed 03/23/2026
Name of Provider or Supplier A Comprehensive Dermatology	Street Address, City, State 1575 W Big Beaver Rd, Suite C12, Troy, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: . . . Based on record review and interview with the laboratory director (LD), the laboratory failed to establish a written policy to assess and monitor staff competency for 14 (February 2024 through March 2026) of 14 months reviewed. Findings include: 1. Record review of the laboratory's procedure manual revealed that the laboratory did not have a written policy for assessing and monitoring staff competency. 2. On 03/23 /2026 at 12:00 pm, an interview with the LD confirmed that the laboratory did not have a competency assessment policy.</p>