

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  23D0956762	<b>(X3) Date Survey Completed</b>  08/31/2021
<b>Name of Provider or Supplier</b>  Dermatology & Skin Surgery Center	<b>Street Address, City, State</b>  300 E Maiden Ln, St Joseph, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The Dermatology & Skin Surgery Center, PLC laboratory was found to be in substantial compliance with CLIA regulations (42 CFR Part 93, effective April 24, 2003). No deficiencies were cited.
<b>D3000</b>	<p><b>FACILITY ADMINISTRATION</b> CFR(s): 493.1100</p> <p>Each laboratory that performs nonwaived testing must meet the applicable requirements under 493.1101 through 493.1105, unless HHS approves a procedure that provides equivalent quality testing as specified in Appendix C of the State Operations Manual (CMS Pub. 7). (a) Reporting of SARS-CoV-2 test results During the Public Health Emergency, as defined in 400.200 of this chapter, each laboratory that performs a test that is intended to detect SARS-CoV-2 or to diagnose a possible case of COVID-19 (hereinafter referred to as a "SARS-CoV-2 test") must report SARS-CoV-2 test results to the Secretary in such form and manner, and at such timing and frequency, as the Secretary may prescribe.</p> <p>This CONDITION is not met as evidenced by:  . Based on record review, lack of documentation, and interview with the Office Administrator, the laboratory failed to report SARS-CoV-2 test results every day of patient testing for 32 of 34 tests performed by the laboratory. Findings include: 1. A tour of the facility on 8/31/21 at 9:01 am revealed the laboratory is performing SARS-CoV-2 testing using the BD Veritor Plus test system. 2. An interview with the Office Administrator on 8/31/21 at 9:32 am revealed the laboratory does not send results of its antigen testing to the health department.</p>