

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D1007218	(X3) Date Survey Completed 10/29/2025
Name of Provider or Supplier Jona K Gill, Md, Pc	Street Address, City, State 14650 E Old Us 12, Chelsea, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the Practice Manager (PM), the laboratory failed to ensure that 6 reagents used for histopathology tissue examination were not available for use beyond their expiration dates. Findings include: 1. On 10/29 /2025 at 9:45 am, the surveyor observed the following expired reagents: a. 1 of 2 bottles of Advantik Red Tissue Marking Dye with an expiration date of 6/30/2024. b. 1 of 2 bottles of Advantik Black Tissue Marking Dye with an expiration date of 6/30 /2024. c. 1 of 2 bottles of Advantik Green Tissue Marking Dye with an expiration date of 6/30/2024. d. 1 of 2 bottles of Advantik Yellow Tissue Marking Dye with an expiration date of 9/30/2024. e. 1 of 2 bottles of Advantik Blue Tissue Marking Dye with an expiration date of 7/31/2024. d. 1 of 5 bottles of Mercedes Scientific Eosin with an expiration date of 7/31/2025. 2. On 10/29/2025 at 9:50 am, an interview with the PM confirmed the reagents were expired.</p>