

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D1033208	(X3) Date Survey Completed 07/23/2019
Name of Provider or Supplier Somerset Skin Centre	Street Address, City, State 255 Kirts Blvd Suite 100, Troy, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5801	<p>TEST REPORT CFR(s): 493.1291(a)</p> <p>The laboratory must have an adequate manual or electronic system(s) in place to ensure test results and other patient-specific data are accurately and reliably sent from the point of data entry (whether interfaced or entered manually) to final report destination, in a timely manner. This includes the following: (a)(1) Results reported from calculated data. (a)(2) Results and patient-specific data electronically reported to network or interfaced systems. (a)(3) Manually transcribed or electronically transmitted results and patient-specific information reported directly or upon receipt from outside referral laboratories, satellite or point-of-care testing locations.</p> <p>This STANDARD is not met as evidenced by: . Based on document review and interview with Testing Personnel #2 (TP2), the office manager, and the Mohs technician, the laboratory failed to have the final mycology potassium hydroxide (KOH) result maintained as part of the patient's electronic medical record (EMR) for 1 (#3) of 40 patient charts audited. Findings include: 1. Document review for 1 (#3) of 40 patient charts audited revealed the final KOH test result was absent in the patient's EMR. 2. During the interview on July 23, 2019 at 1:00 PM with TP2, the office manager, and the Mohs technician confirmed the test was ordered, but no results were documented in the EMR.</p>