

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D1033208	(X3) Date Survey Completed 02/25/2025
Name of Provider or Supplier Somerset Skin Centre	Street Address, City, State 255 Kirts Blvd Suite 100, Troy, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: . Based on observation and interview with Testing Personnel (TP1), the laboratory failed to ensure 4 bottles of reagents were not used beyond their expiration date. Findings include: 1. The surveyor observed the following 4 bottles of reagents available for patient testing on 2/25/2025 at 9:30 am: a. 1 bottle of Mercedes Scientific Blue Tissue Marking Dye with an expiration date of 11/30/2024 b. 1 bottle of Mercedes Scientific Red Tissue Marking Dye with an expiration date of 1/31/2025 c. 1 bottle of 10% KOH Reagent with an expiration date of 1/11/2025 d. 1 bottle of Platinumline Eosin Y Stain with an expiration date of 09/27/2024 2. An interview conducted on 02/25/2025 at 9:45 am with the TP1 confirmed the reagents were expired and available for use.</p>