

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D1034340	(X3) Date Survey Completed 05/28/2026
Name of Provider or Supplier Pediatric Consultants Of Troy	Street Address, City, State 50720 Schoenherr Road, Shelby Township, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2006	TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b) (b)The laboratory must examine or test, as applicable, the proficiency testing samples it receives from the proficiency testing program in the same manner as it tests patient specimens. This testing must be conducted in conformance with paragraph (b)(4) of this section. If the laboratory's patient specimen testing procedures would normally require reflex, distributive, or confirmatory testing at another laboratory, the laboratory should test the proficiency testing sample as it would a patient specimen up until the point it would refer a patient specimen to a second laboratory for any form of further testing. This STANDARD is not met as evidenced by: . Based on record review and interview with the technical consultant (TC), the laboratory failed to ensure proficiency testing samples were tested in the same manner as patient specimens for 5 (2024 Event 3; 2025 Events 1, 2, and 3; and 2026 Event 1) of 6 proficiency testing events reviewed. Findings include: 1. A record review of proficiency testing records revealed proficiency testing samples were tested twice during Event 3 of 2024 and Events 1, 2, and 3 of 2025, and Event 1 of 2026. 2. An interview with the TC on 05/28/2026 at 12:30 pm confirmed proficiency testing samples were tested twice for the events listed above. 3. On 05/28/2026 at 12:32 pm, the surveyor requested a copy of the laboratory's Proficiency Testing policy. Record review of the laboratory's Proficiency Testing policy, updated in March 2026, revealed proficiency testing samples are to be tested once. 4. An interview with the TC on 05/28/2026 at 12:35 pm revealed the policy was updated following a citation issued at another CLIA certified laboratory because the previous policy instructed personnel to test proficiency testing samples in duplicate.