

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  23D1069404	<b>(X3) Date Survey Completed</b>  04/20/2022
<b>Name of Provider or Supplier</b>  98 Point 6 Emergicenter	<b>Street Address, City, State</b>  1540 Lake Lansing Road Suite 203, Lansing, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2016</b>	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: . Based on review of the CMS database and review of the American Proficiency Institute (API) final proficiency testing reports, it was determined the laboratory failed to successfully participate in a CMS approved proficiency testing program for the specialty of hematology for 2 (1st event in 2022 and 3rd event 2021) of 3 testing events in 2021 and 2022. Findings include: Review of the CMS database and the API proficiency testing reports showed unsatisfactory performance for 2 of 3 consecutive proficiency testing events for the specialty of hematology. Refer to D2131. .</p>
<b>D2131</b>	<b>HEMATOLOGY</b>

CFR(s): 493.851(g)

Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

. Based on review of the CMS database and review of the American Proficiency Institute (API) final proficiency testing reports, it was determined the laboratory failed to successfully participate in the specialty of hematology for 2 (3rd event of 2021 and the 1st event of 2022) of 3 testing events. Findings include: 1. A review of the CMS database and review of the API final proficiency testing reports revealed the laboratory attained the following for the specialty of hematology proficiency testing scores: Specialty of Hematology PT Event Score 3rd of 2021 66 1st of 2022 0%