

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D1091053	(X3) Date Survey Completed 08/01/2019
Name of Provider or Supplier Feinstein Laboratory	Street Address, City, State 29829 Telegraph Road Ste 203, Southfield, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the Laboratory Director (LD), the laboratory failed to establish written procedures for the testing of sodium, potassium, and chloride for 2 (August 2017 to August 2019) of 2 years. Findings include: 1. A record review of chemistry testing procedures exposed a lack of written procedures for sodium, potassium, and chloride testing using the Mindray BS 200 chemistry analyzer. 2. When requested on 8/1/19 at 9:55 am, the LD failed to provide established procedures for sodium, potassium, and chloride testing. 3. An interview on 8/1/19 at 9:55 am with the LD confirmed the laboratory did not establish written procedures for sodium, potassium, and chloride testing.</p>