

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 23D1092620	<b>(X3) Date Survey Completed</b> 05/19/2022
<b>Name of Provider or Supplier</b> Physical Medicine Rehabilitation And Consultant	<b>Street Address, City, State</b> 21675 Coolidge Hwy Suite A, Oak Park, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5401</b>	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the Technical Consultant (TC), the laboratory failed to have a written procedure manual available to laboratory personnel for 1 (April 2022 to May 2022) of 1 month reviewed. Findings include: 1. A review of the laboratory's documentation on 5/19/22 revealed a lack of procedure manual. 2. An interview on 5/19/22 at 9:50 am with the TC revealed the procedure manual was at their office during the creation of the laboratory's allegation of compliance and was not present at the laboratory.</p>