

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  23D2001422	<b>(X3) Date Survey Completed</b>  12/06/2023
<b>Name of Provider or Supplier</b>  Jaro Company Llc	<b>Street Address, City, State</b>  1349 S Rochester Road Suite 210, Rochester Hills, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by:                      . Based on observation and interview with Testing Personnel (TP) #1, the laboratory failed to calibrate and/or replace the expired thermometer in the RCA refrigerator for 20 (April 2022 to December 2023) of 20 months in use. Findings include: 1. During a tour of the laboratory on 12/06/2023 at 9:39 am, the surveyor observed a Cardinal Health thermometer in the RCA refrigerator with an expiration date of 02/16/2007. 2. An interview on 12/06/2023 at 9:39 am, TP1 confirmed the thermometer in the RCA refrigerator had expired and was not calibrated and/or replaced.</p>