

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D2062073	(X3) Date Survey Completed 03/10/2025
Name of Provider or Supplier Forefront Dermatology, S C	Street Address, City, State 2525 East Paris Avenue S E Suite 100, Grand Rapids, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3013	<p>FACILITIES CFR(s): 493.1101(e)</p> <p>Records and, as applicable, slides, blocks, and tissues must be maintained and stored under conditions that ensure proper preservation.</p> <p>This STANDARD is not met as evidenced by: . Based on observation and interview with dermatology personnel #1 and #2, the laboratory failed to ensure histopathology tissue slides were coverslipped to ensure tissue preservation for nine (Patients 1-9) of 10 patient cases reviewed. Findings include: 1. A review of 10 patient cases revealed the following nine cases had slides with significant bubbling over the stained tissue sections: a. Patient 1 had histopathology testing performed on 02/13/2025. b. Patient 2 had histopathology testing performed on 12/04/2024. c. Patient 3 had histopathology testing performed on 09/19/2024. d. Patient 4 had histopathology testing performed on 09/19/2024. e. Patient 5 had histopathology testing performed on 06/20/2024. f. Patient 6 had histopathology testing performed on 04/17/2024. g. Patient 7 had histopathology testing performed on 01/25/2024. h. Patient 8 had histopathology testing performed on 10/10/2023. i. Patient 9 had histopathology testing performed on 04/18/2023. 2. An interview on 3/10/25 at 2:00 pm with dermatology personnel #1 and #2 confirmed the slides had significant bubbling over tissue sections.</p>