

| | | |
|--|---|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 23D2065421 | (X3) Date Survey Completed 01/25/2022 |
| Name of Provider or Supplier Comprehensive Pain Solutions, Pllc | Street Address, City, State 37459 Schoolcraft Road, Livonia, MI | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D5417 | <p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: . Based on observation, record review, and interview with the Laboratory Director (LD), the laboratory failed to ensure control materials did not exceed their expiration date for the negative control in use. Findings include: 1. The surveyor observed a clear plastic bottle in the refrigerator labeled "Surine" with the expiration date of "1/21" on 1/25/22 at 11:06 am. 2. A review of the laboratory's "Reagent Handling" policy revealed a section stating, "All reagents, calibrators, controls and solutions are used within their indicated expiration date, whether this date is manufacturer-provided for purchased materials or assigned for laboratory prepared materials." 3. An interview on 1/25/22 at 11:06 am with the Laboratory Director revealed the laboratory was currently using the expired solution as a negative control in toxicology testing and confirmed it was expired.</p> |