

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D2077425	(X3) Date Survey Completed 05/15/2024
Name of Provider or Supplier Dermatology Specialists Of Brighton	Street Address, City, State 2305 Genoa Business Park Dr, Suite 180, Brighton, MI	
<p>For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.</p>		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: . Based on observation, interview and record review, the laboratory failed to ensure the Eosin Y Alcoholic 25% Histopathology Stain, was not used beyond the expiration date for 11 months (expiration date 6/30/23). Findings include: 1. The surveyor observed 1 bottle of Eosin Y Alcoholic 25% Histopathology Stain with the expiration date 6/30/23 in the Flammable Storage Cabinet on 05/15/24 at 8:08 am. 2. An interview of the laboratory manager was conducted on 5/15/2024 at 8:10 am and confirmed the stain was expired. 3. A record review of the laboratory's "Regulatory Overview" policy indicated a section titled, "Reagents, Materials and Supplies" which stated "No reagent, solution, mediates or materials should be used after their expiration date."</p>