

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D2077429	(X3) Date Survey Completed 02/05/2024
Name of Provider or Supplier Dermatology Specialists Of Shelby	Street Address, City, State 50505 Schoenherr Road, Ste 325, Shelby Township, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: . Based on observation and interview with the Assistant Office Manager, the laboratory failed to ensure its Frozen Embedding Media had not exceeded its expiration dates for 4 of 14 total bottles observed. Findings include: 1. The surveyor observed the 14 bottles of laboratory's POLARSTAT Frozen Embedding Media on 2/5/24 at 11:32 pm the following had exceeded expiration dates: a. 1 bottle with the lot number 118028 and the expiration date 1/31/23. b. 3 bottles with the lot number 123390 and the expiration date 5/31/23. 2. An interview on 2/5/24 at 11:38 am with the Assistant Office Manager confirmed the Frozen Embedding Media had exceeded expiration dates for the bottles listed above.</p>