

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D2100814	(X3) Date Survey Completed 08/27/2024
Name of Provider or Supplier Adams Medical Center	Street Address, City, State 23850 Van Born Road, Dearborn Heights, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An unannounced revisit survey was performed on August 27, 2024 as a follow-up to the July 22, 2024 recertification survey. The Department of Licensing and Regulatory Affairs has evaluated this facility and determined it is not compliance with CLIA regulations at the standard level (42 CFR Part 93, effective April 24, 2003).
D5433	<p> MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(b)(1) </p> <p> For equipment, instruments, or test systems developed in-house, commercially available and modified by the laboratory, or maintenance and function check protocols are not provided by the manufacturer, the laboratory must establish a maintenance protocol that ensures equipment, instrument, and test system performance that is necessary for accurate and reliable test results and test result reporting. The laboratory must perform and document the maintenance activities specified in paragraph (b)(1)(i) of this section. </p> <p> This STANDARD is not met as evidenced by: . Based on observation and interview with the Testing Personnel, the laboratory failed to establish maintenance protocols to ensure centrifuge performance for accurate and reliable qualitative urine toxicology results for one (August 2024) of one month the laboratory started using the centrifuge. Findings include: 1. The surveyor observed the laboratory's centrifuge used to prepare urine specimens for qualitative urine toxicology testing on 8/27/24 at 9:36 am. 2. An interview with the Testing Personnel at 9:36 am revealed the laboratory did establish a policy for centrifuge maintenance and no maintenance or function checks for the centrifuge were performed prior to use for patient testing. </p>