

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 23D2153556	<b>(X3) Date Survey Completed</b> 06/03/2021
<b>Name of Provider or Supplier</b> Hamzavi Derm Pathology Bloomfield Satellite	<b>Street Address, City, State</b> 5216 Rector Court, Bloomfield Hills, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5893</b>	<p>POSTANALYTIC SYSTEMS QUALITY ASSESSMENT CFR(s): 493.1299(b)(c)</p> <p>(b) The postanalytic systems quality assessment must include a review of the effectiveness of corrective actions taken to resolve problems, revision of policies and procedures necessary to prevent recurrence of problems, and discussion of postanalytic systems quality assessment reviews with appropriate staff. (c) The laboratory must document all postanalytic systems quality assessment activities.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the Laboratory Director, the laboratory failed to document postanalytic systems quality assessment activities for 2 (June 2019 to June 2021) of 2 years reviewed. Findings include: 1. A review of the laboratory's quality assessment records revealed a lack of postanalytic systems quality assessments. 2. A review of the laboratory's "Quality Assurance Manual" revealed a section stating, "Each of the laboratory's quality systems will undergo assessment on a regular basis to maintain and improve laboratory performance and services." 3. An interview on 6/3/21 at 10:22 am with the Laboratory Director revealed the laboratory did not document postanalytic systems quality assessment activities.</p>