

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D2193424	(X3) Date Survey Completed 02/24/2021
Name of Provider or Supplier Warren Care Pharmacy	Street Address, City, State 11460 E 12 Mile Road, Warren, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D1002	<p>REPORTING OF SARS-CoV-2 TEST RESULTS</p> <p>During the Public Health Emergency, as defined in 400.200 of this chapter, each laboratory that performs a test that is intended to detect SARS-CoV-2 or to diagnose a possible case of COVID-19 (hereinafter referred to as a "SARS-CoV-2 test") must report SARS-CoV-2 test results to the Secretary in such form and manner, and at such timing and frequency, as the Secretary may prescribe.</p> <p>This CONDITION is not met as evidenced by: . Based on record review and interview with the Laboratory Director, the laboratory failed to report SARS-Co-V-2 test results as required for 3 of 3 tests performed since the laboratory has been in operation. Findings include: 1. The surveyor requested the facility's SARS-CoV-2 testing and reporting policy via email correspondence on 2/23/21 at 9:00 am and it was not made available. 2. An interview with the Laboratory Director on 2/24/21 at 3:00 pm revealed the laboratory had performed a total of three patient tests with the BD Veritor Plus SARS-CoV-2 antigen test system and it did not report those results to the health department. 3. The Laboratory Director confirmed the findings on 2/24/21 at 3:00 pm.</p>