

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 23D2215148	(X3) Date Survey Completed 05/20/2021
Name of Provider or Supplier Anthology Of Troy	Street Address, City, State 3400 Livernois, Troy, MI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D1002	<p>REPORTING OF SARS-CoV-2 TEST RESULTS</p> <p>During the Public Health Emergency, as defined in 400.200 of this chapter, each laboratory that performs a test that is intended to detect SARS-CoV-2 or to diagnose a possible case of COVID-19 (hereinafter referred to as a "SARS-CoV-2 test") must report SARS-CoV-2 test results to the Secretary in such form and manner, and at such timing and frequency, as the Secretary may prescribe.</p> <p>This CONDITION is not met as evidenced by:</p> <p>. Based on record review and interview with the Laboratory Director (LD), the laboratory failed to report SARS-Co-V-2 test results as required for 7 (11/16/2020 to 5/20/2021) of 7 months of testing. Findings include: 1. A record review of the "Anthology Senior Living Covid-19 Policy & Procedure" revealed in "Reporting Test Results to State Governing Entities: Depending on the state specific regulations, communities are required to report either their data for all testing's completed for each individual test or for all test that result in a positive result. Communities will comply with their specific state regulations requiring reporting." 2. When queried during the phone interview on 5/20/2021 at 2:17 pm, the surveyor was informed that for 7 (11/16/2020 to 5/20/2021) of 7 months of testing only positive testing results are reported to the Oakland County Health Department. 3. An interview on 5/20/2021 at 2:17 pm, the LD confirmed the laboratory was not reporting negative and positive patient SARS-CoV-2 testing to the local health department.</p>