

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 23D2274436 | (X3) Date Survey Completed 08/20/2025 |
| Name of Provider or Supplier Zarka Dermatology | Street Address, City, State 1057 Suncrest Drive, Lapeer, MI | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D3011 | <p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: . Based on observation and interview with the Mohs Technician, the laboratory failed to ensure staff protection from chemically hazardous materials for 21 (December 2023 to August 2025) of 21 months reviewed. Findings include: 1. An observation of the laboratory on 08/10/2025 at 9:30 am revealed a lack of eye wash station. 2. An interview on 08/10/2025 at 9:30 am with the Mohs Technician confirmed the laboratory removed the eyewash station for remodel and failed to reinstall it. ***This is a repeat deficiency***</p> |
| D5217 | <p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on record review and interview with the Laboratory Director (LD), the laboratory failed to verify the accuracy of its testing at least twice annually for Histopathology testing in 1 (1 biannual period of 2024) of 4 biannual periods reviewed. Findings include: 1. A review of the laboratory's twice annual verification of accuracy documentation revealed lack of documentation of twice-annual verification of accuracy for Histopathology testing in 2024. 2. On 08/10/2025 at 10:15</p> |

am, a request was made to the LD to provide verification of accuracy documentation for 2024. 3. An interview conducted on 08/10/2025 at 10:15 am with the LD confirmed that verification of accuracy testing for Histopathology testing had not been performed twice in 2024.