

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  24D0405305	<b>(X3) Date Survey Completed</b>  11/08/2022
<b>Name of Provider or Supplier</b>  Murray County Medical Center	<b>Street Address, City, State</b>  2042 Juniper Avenue, Slayton, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5775</b>	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c)</p> <p>(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. (c) The laboratory must document all test result comparison activities.</p> <p>This STANDARD is not met as evidenced by: . Based on observation, document review and interview with laboratory personnel, the laboratory failed to evaluate the relationship between test results obtained from the Hematology analyzer and a manual testing method at least twice annually. Findings are as follows: 1. The laboratory performed Hematology testing as confirmed by the General Supervisor (GS) during a tour of the laboratory on 11/08/22, at 12:05 p.m. 2. A Horiba ABX XL80 hematology analyzer was observed as present and available for use during the tour. The GS indicated the laboratory performed and reported automated and manual White Blood Cell differential testing. 3. The following procedures located in the Murray County Medical Center Policy &amp; Procedures manual: - Hematology Pentra XL80 CBC with 5 Part Automated Differential - Manual Differential - Back-up Method Comparison did not include a requirement to compare automated and manual differential testing twice annually. Twice annual comparisons of these test methods was not found in laboratory records. The laboratory was unable to provide comparison records upon request. 4. In an interview on 11/09/22, at 10:30 a.m., the GS confirmed twice annual comparisons of the automated and manual White Blood Cell differential testing had not been performed. .</p>