

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D0406017	(X3) Date Survey Completed 01/05/2021
Name of Provider or Supplier Centracare Paynesville	Street Address, City, State 200 1st Street W, Paynesville, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on document review and interview with laboratory personnel, the laboratory failed to investigate an unacceptable Chemistry proficiency testing (PT) result for 1 analyte in 2019. Findings are as follows: 1. The laboratory performed Chemistry testing as confirmed by the Technical Supervisor (TS) during the entrance interview at 8:05 a.m. on 01/05/21. 2. The laboratory performed PT using the Wisconsin State Laboratory of Hygiene (WSLH) and American Proficiency Institute programs. 3. The laboratory received an unacceptable pH PT result in 1 of the 21 scheduled PT events completed in 2019. See below. WSLH 2019 Blood Gas 3rd event Sample Test Lab result WSLH range BG-11 pH 7.125 7.168-7.248 4. Investigation of unacceptable PT results was required as established in the Proficiency Testing Policy located in the laboratory's electronic document management software PolicyStat. 5. An investigation of the unacceptable PT result was not found during review of laboratory records. The laboratory was unable to provide investigation documentation upon request. 6. In an interview at 11:00 a.m. on 01/05/21, the TS confirmed the above finding.</p>