Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D0680316	(X3) Date Survey Completed 05/20/2022
Name of Provider or Supplier Associated Skin Care Specialists Pa	**Street Address, City, State** 18315 Cascade Dr Suite #150, Eden Prairie, MN	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: . Based on document review and interview with laboratory personnel, the laboratory failed to perform and document testing accuracy verification activities at least twice annually for one of three microscopic examinations in 2020 and two of three microscopic examinations in 2021. Findings are as follows: 1. The laboratory performed microscopic examinations for fungus, parasites, and viruses under the specialty of Microbiology as confirmed by the Laboratory Specialist (LS) during a tour of the laboratory at 10:15 a.m. on 05/20/22. 2. Requirements for twice annual verification of accuracy testing for Scabies parasite preparations (S), and Tzanck virus preparations (T) were established in the laboratory's Proficiency Testing for Provider Performed Microscopic Examinations procedure found in the KOH, Tzanck, Scabies, Pregnancy UA manual. 3. One verification of testing accuracy document for Scabies preparations was found during review of 2020 laboratory records. The laboratory was unable to provide a second accuracy verification from 2020 for Scabies preparations upon request. 4. Verification of testing accuracy documents for Scabies and Tzanck preparations were not found during review of 2021 laboratory records. The laboratory was unable to provide accuracy verification documents from 2021 for Scabies and Tzanck preparations upon request. 5. The laboratory's Scabies test log indicated patient samples received microscopic examinations for parasites in 2020 and 2021. The Tzanck test log indicated a single patient sample received microscopic examination for viruses in 2021. See below. 2020 Test Number of patients S 1 2021 Test Number of patients S 1 T 1 6. In an interview at 12:25 p.m. on 05/20/22, the LS confirmed the above finding.