

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D0882688	(X3) Date Survey Completed 09/24/2020
Name of Provider or Supplier Scientific Research Consortium, Inc	Street Address, City, State 2840 Patton Road, Saint Paul, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6127	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: . Based on document review and interview with laboratory personnel, the Technical Supervisor failed to ensure 4 of 5 testing personnel received a competency evaluation at least semiannually during the first year of patient specimen testing. Findings are as follows: 1. The laboratory performed Amino Acid and Fatty Acid Analysis under the subspecialty of Chemistry as confirmed by Testing Supervisor 2 (TS2) and Testing Supervisor 3 (TS3) during the entrance interview and tour of the lab at 10:05 a.m. on 09/24/20. 2. Testing Personnel 3 (TP3), Testing Personnel 4 (TP4), Testing Personnel 5 (TP5) and Testing Personnel 6 (TP6) were listed on the Laboratory Personal Report (CLIA) Form CMS-209 as high complexity testing personnel. 3. Semiannual competency assessment were not found for 4 of 5 testing personnel (TP3, TP4, TP5, TP6) during review of laboratory personnel records. The laboratory was unable to provide a semiannual competency assessment for the testing personnel upon request. 5. In an interview at 9:15 a.m. on 09/24/20, TS3 confirmed the above findings. .</p>