

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D0957471	(X3) Date Survey Completed 04/16/2026
Name of Provider or Supplier Dermatology Consultants, Pa	Street Address, City, State 576 Bielenberg Drive Suite #200, Woodbury, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The Dermatology Consultants, PA laboratory was found to be out of compliance with the regulations of the Clinical Laboratory Improvement Amendments of 1988 (42 C.F.R. part 493) upon completion of the recertification survey performed on April 16, 2026. The following standard-level deficiencies were cited: 493.1254 Maintenance and function checks .
D5433	<p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(b)(1)</p> <p>(b)(1)(i) Establish a maintenance protocol that ensures equipment, instrument, and test system performance that is necessary for accurate and reliable test results and test result reporting. (b)(1)(ii) Perform and document the maintenance activities specified in paragraph b(1)(i) of this section.</p> <p>This STANDARD is not met as evidenced by: . Based on observation, document review, and interview with laboratory personnel, the laboratory failed to perform and document routine Microscopy equipment maintenance according to established procedures in 2024. Findings are as follows: 1. The laboratory performed Mohs Micrographic surgery under the Histopathology specialty and Mycology and Parasitology testing under the Microbiology specialty as confirmed by the Clinic Manager during a tour of the laboratory at 10:08 a.m. on 4/16 /26. 2. The following items were observed as present and available for use during the tour: A Leica DM 1000 microscope used for Mohs Micrographic Surgery slide review A Nikon Alphaphot-2 microscope used for Potassium Hydroxide (KOH) fungal and Scabies testing 3. The laboratory was required to perform monthly microscope maintenance as defined in the Microscope Use policy found in the Procedure Log binder provided by the laboratory and reviewed on the date of survey. 4. Performance of microscope maintenance was not documented on the Monthly Microscope Maintenance Log for seven of twelve months in 2024. See below. Months missing maintenance for Leica and Nikon microscopes in 2024: April June July September</p>

October November December 5. In an interview at 12:10 p.m., the Clinic Manager confirmed the above findings. .