

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  24D0969699	<b>(X3) Date Survey Completed</b>  11/01/2024
<b>Name of Provider or Supplier</b>  Skin Care Doctors Pa	<b>Street Address, City, State</b>  14000 Nicollet Ave South Suite 304, Burnsville, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	<p>. The Skin Care Doctors PA laboratory was found to be out of compliance with the regulations of the Clinical Laboratory Improvement Amendments of 1988 (42 C.F.R. part 493) upon completion of the initial survey performed on November 1, 2024. The following standard-level deficiency was cited: 493.1291 Test report .</p>
<b>D5805</b>	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: . Based on observation, document review, and interview with laboratory personnel, the laboratory failed to ensure one of two test reports reviewed on date of survey included complete test results in 2024. Findings are as follows: 1. The laboratory performed Mohs micrographic surgery which included tissue inking and Histopathology slide interpretation as confirmed by the Practice Administrator (PA) during a tour of the laboratory at 10:06 a.m. on 11/01/24. The laboratory began testing on May 1, 2024. 2. The following equipment was observed as present and available for use during the tour: Leica Autostainer XL Leica ST4020 autostainer Leica CV5030 coverslipper Leica CM1850 cryostat (2) Olympus CX41 microscope 3. Mohs case 24-091, performed on August 22, 2024, was one of two cases reviewed on date of survey. The Mohs map, the Skin Care Doctors Mohs log, and associated patient slides indicated the surgery was performed in two stages. The Mohs map</p>

documentation indicated the tissue margins were clear of tumor after the second surgical stage. 4. The operative report (test report) included a description of the first surgical stage. The report did not include results of the first stage nor a description indicating the second and final result of the surgery. 5. In an interview at 12:35 p.m. on 11/01/24, the PA confirmed the above finding.