

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D0992852	(X3) Date Survey Completed 03/28/2022
Name of Provider or Supplier Centracare Clinic- Plaza Dermatology	Street Address, City, State 1900 Centracare Circle Suite 2575, Saint Cloud, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on document review and interview with laboratory personnel, the laboratory failed to perform and document testing accuracy verification activities for 3 of 3 microscopic examinations at least twice annually in 2021. Findings are as follows: 1. The laboratory performed microscopic examinations for fungus, parasites, and viruses under the specialty of Microbiology as confirmed by the Dermatology Manager (DM) during a tour of the laboratory at 10:00 a.m. on 03/28/22. 2. Requirements for twice annual verification of accuracy testing for KOH fungal preparations (KOH), Scabies parasite preparations (S), and Tzanck virus preparations (T) were established in the laboratory's Dermatology Quality Assurance Plan found in the CLIA Quality Control manual. 3. Verification of testing accuracy documentation for the three microscopic examinations was not found during review of laboratory records from 2021. The laboratory was unable to provide the missing documentation upon request. 4. The laboratory's KOH, Tzanks (sic) and Scabies log indicated patient samples received microscopic examinations for fungus and parasites in 2021. See below. 2021 Test Number of patients KOH 66 S 5 T 0 5. In an interview at 11:30 a.m. on 03/28/22, the DM confirmed the above finding.</p>