

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 24D1014635	<b>(X3) Date Survey Completed</b> 06/18/2018
<b>Name of Provider or Supplier</b> Skin Care Drs Pa	<b>Street Address, City, State</b> 1350 Lesauk Drive, Sartell, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on document review and interview with laboratory personnel, the laboratory failed to perform and document activities used to verify the accuracy of Microbiology test procedures at least twice annually in 2017. Findings are as follows: 1. The laboratory performed Microscopic Examinations for fungus and parasites under the specialty of Microbiology as confirmed by the Histotechnician (HT) during a tour of the laboratory on 06/18/18 at 10:05 a.m. 2. Laboratory records indicated 5 fungal examinations and 1 parasite examination had been performed in 2017. 3. Documentation of 2017 Microscopic Examination verification of accuracy was not found in laboratory records. The laboratory was unable to provide verification of accuracy documentation from 2017 upon request. 4. In an interview on 06/18/18 at 12:00 p.m., the HT confirmed Microscopic Examination accuracy had not been verified in 2017.</p>
<b>D5401</b>	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by:</p>

. Based on document review and interview with laboratory personnel, the laboratory failed to include a procedure for microscopic examinations in the procedure manual. Findings are as follows: 1. The laboratory performed Microscopic Examinations for parasites under the specialty of Microbiology as confirmed by the Histotechnician (HT) during a tour of the laboratory on 06/18/18 at 10:05 a.m. 2. A written procedure for the microscopic examination of parasites was not found during review of the Skin care Doctors Mohs Lab Sartell Policy and Procedure manual. The laboratory staff was unable to provide a procedure upon request. 3. In an interview on 06/18/18 at 12:05 p. m., the HT confirmed the above finding.