Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 24D1016491	**(X3) Date Survey Completed** 05/13/2025
Name of Provider or Supplier Mankato Clinic At Daniels Health Center	**Street Address, City, State** 1901 Old Minnesota Avenue, Saint Peter, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	The Mankato Clinic at Daniels Health Center laboratory was found to be out of compliance with the regulations of the Clinical Laboratory Improvement Amendments of 1988 (42 C.F.R. part 493) upon completion of the recertifiication survey performed on May 13, 2025. The following standard-level deficiencies were cited: 493.1291 Test report
D5807	TEST REPORT CFR(s): 493.1291(d) (d) Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results. This STANDARD is not met as evidenced by: . Based on observation, document review, and interview with the laboratory personnel, the laboratory failed to ensure four of fifteen Urinalysis reference intervals were available to the authorized person responsible for using the test results. Findings are as follows: 1. The laboratory performed urinalysis testing as confirmed by Testing Personnel 1 during a tour of the laboratory at 10:06 a.m. on 5/13/25. 2. A Nikon Eclipse E2000 microscope was observed as present and available for use for urine microscopic examination during the tour. 3. Urine microscopic examination reference intervals were not established on a patient test report from 5/15/23. See below: Missing reference intervals: White blood cells /hpf Red blood cells /hpf Epithelial Cells /hpf Bacteria /hpf 4. In an interview at 1:09 p.m. on 5/13/25, the Technical Consultant (TC) confirmed the above findings. 5. The laboratory performed 1,131 urine microscopic exams since March 2023 as indicated by the TC in an email received at 12:42 p.m. on 5/15/25.