

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  24D1085475	<b>(X3) Date Survey Completed</b>  02/28/2018
<b>Name of Provider or Supplier</b>  Associated Skin Care Specialists Pa - Maple Grove	<b>Street Address, City, State</b>  9600 Upland Lane N #250, Maple Grove, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by:                      . Based on observation and interview with laboratory personnel, the laboratory failed to ensure Histopathology testing material was not used after the expiration date had been exceeded. Findings are as follows: 1. The laboratory performed Mohs Micrographic Surgery under the subspecialty Histopathology as confirmed by the Laboratory Specialist (LS) during a tour of the laboratory on 02/28/18 at 1:05 p.m. 2. One bottle of Mounting Medium, lot 350525 and expiration date 10/2017, was observed as present and available for use during the tour of the laboratory. 3. In an interview on 02/28/18 at 1:06 p.m., the LS confirmed the Mounting Medium was used after the expiration date had been exceeded.</p>