

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D2047551	(X3) Date Survey Completed 10/11/2018
Name of Provider or Supplier Skincare Drs Pa	Street Address, City, State 7373 France Ave So Suite 304, Edina, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3031	<p>RETENTION REQUIREMENTS CFR(s): 493.1105(a)(3)</p> <p>Analytic systems records. Retain quality control and patient test records (including instrument printouts, if applicable) and records documenting all analytic systems activities specified in 493.1252 through 493.1289 for at least 2 years.</p> <p>This STANDARD is not met as evidenced by:</p> <p>. Based on document review and interview with laboratory personnel, the laboratory failed to retain Mohs Micrographic Surgery quality control (QC) slides for at least 2 years. The laboratory performed approximately 350 Mohs Micrographic Surgery procedures annually. Findings are as follows: 1. The laboratory performed Mohs Micrographic Surgery as confirmed by the Histotechnician (HT) during a tour of the laboratory on 10/11/18 at 1:05 p.m. 2. During the tour, the HT stated separate Mohs QC slides were created using excess tissue from the first patient case of the day. 3. Mohs QC slides associated with cases 17-033 from 05/15/17 and 18-065 from 05/14/18 were not available on date of survey. The laboratory was unable to provide the Mohs QC slides upon request. 4. In an interview on 10/11/18 at 1:20 p.m., the HT indicated the QC slides were discarded after surgeon review on each day of patient testing.</p>